



California Medical Device Recall Information



Recall Name

GE Healthcare Recalls

Engström Ventilator and Aespire View, Aisys, and Avance Anesthesia Machines Due To a Potential Safety Issue

| Recall Date | Product Description | Recalling Firm | Recall Reason |
|--------------|---|---|---|
| 10/16/13 | <ul style="list-style-type: none">• Engström Ventilator• Aespire View Anesthesia Machines• Aisys Anesthesia Machines• Avance Anesthesia Machines | GE Healthcare, LLC. Madison, WI | <i>Unresponsive buttons on the display of Engström Ventilator and Aespire View, Aisys, and Avance Anesthesia Machines. This may result in inability to access menu functions and delays in treatment.</i> |
| Recall Class | Product Identification | Distribution | Affected Dates |
| I | Follow the link below to view: Affected Product Serial Numbers | CA , nationwide | Manufactured between: April 23, 2013 and July 22, 2013. |

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm371091.htm>